

MC-21-J电阻箱出厂检测报告					
生产厂家：泸州泸电科技有限公司			官网：http://ldkj.ywtva.com		
机身编号：	MC-21-J 20231124025		测试环境：	温度：7° 湿度：88	
测试日期：	2025年6月8日		测试人：	江明友	
测试设备：	三杰 SJ2515P		全档0位阻值：	0.135 Ω	
说明： 所有单位均为 Ω 实测值为0位阻值归零后测试设备实显值，已除掉0位阻值			测试结果：		
标称值	标称误差	合格范围	实测值	判断结果	备注
0.1	20%	0.08~0.12	0.112	PASS	
0.2	20%	0.16~0.24	0.208	PASS	
0.3	20%	0.24~0.36	0.319	PASS	
0.4	20%	0.32~0.48	0.424	PASS	
0.5	20%	0.4~0.6	0.525	PASS	
0.6	20%	0.48~0.72	0.621	PASS	
0.7	20%	0.56~0.84	0.732	PASS	
0.8	20%	0.64~0.96	0.822	PASS	
0.9	20%	0.72~1.08	0.933	PASS	
1	5%	0.95~1.05	1.006	PASS	
2	5%	1.9~2.1	2.007	PASS	
3	5%	2.85~3.15	2.996	PASS	
4	5%	3.8~4.2	3.998	PASS	
5	5%	4.75~5.25	4.991	PASS	
6	5%	5.7~6.3	5.992	PASS	
7	5%	6.65~7.35	6.987	PASS	
8	5%	7.6~8.4	8.027	PASS	
9	5%	8.55~9.45	9.007	PASS	
10	2%	9.8~10.2	10.019	PASS	
20	2%	19.6~20.4	20.058	PASS	
30	2%	29.4~30.6	30.029	PASS	
40	2%	39.2~40.8	40.041	PASS	
50	2%	49~51	50.048	PASS	
60	2%	58.8~61.2	59.989	PASS	
70	2%	68.6~71.4	70.098	PASS	
80	2%	78.4~81.6	80.084	PASS	
90	2%	88.2~91.8	90.036	PASS	
100	2%	98~102	99.94	PASS	
200	2%	196~204	200.09	PASS	
300	2%	294~306	300.04	PASS	
400	2%	392~408	400.02	PASS	
500	2%	490~510	499.76	PASS	
600	2%	588~612	600.42	PASS	
700	2%	686~714	700.25	PASS	
800	2%	784~816	799.82	PASS	

900	2%	882~918	898.79	PASS	
1000	2%	980~1020	999	PASS	
2000	2%	1960~2040	1999	PASS	
3000	2%	2940~3060	2999	PASS	
4000	2%	3920~4080	3997	PASS	
5000	2%	4900~5100	4999	PASS	
6000	2%	5880~6120	5995	PASS	
7000	2%	6860~7140	6994	PASS	
8000	2%	7840~8160	8001	PASS	
9000	2%	8820~9180	9002	PASS	
10000	2%	9800~10200	10003	PASS	
20000	2%	19600~20400	19987	PASS	
30000	2%	29400~30600	30002	PASS	
40000	2%	39200~40800	39954	PASS	
50000	2%	49000~51000	49971	PASS	
60000	2%	58800~61200	59938	PASS	
70000	2%	68600~71400	69928	PASS	
80000	2%	78400~81600	80110	PASS	
90000	2%	88200~91800	89853	PASS	
100000	2%	98000~102000	100051	PASS	
200000	2%	196000~204000	199880	PASS	
300000	2%	294000~306000	300370	PASS	
400000	2%	392000~408000	400110	PASS	
500000	2%	490000~510000	499830	PASS	
600000	2%	588000~612000	598850	PASS	
700000	2%	686000~714000	700470	PASS	
800000	2%	784000~816000	800360	PASS	
900000	2%	882000~918000	903440	PASS	